

Tapered HD Drill Usage Chart

Description	Part Number	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	
2.0mm HD Drill	TSD2020HD																					
2.0mm HD Drill, Extended Shank	TSD4020HD																					
2.5mm HD Drill	TSD2025HD																					
2.5mm Depth Drill, 7.5mm Stop	TSD202507																					
2.5mm Depth Drill, 9mm Stop	TSD202509																					
2.5mm Depth Drill, 10.5mm Stop	TSD202510																					
2.5mm Depth Drill, 12mm Stop	TSD202512																					
2.5mm Depth Drill, 15mm Stop	TSD202515																					
3.0mm HD Crestal Bone Drill	TSC2030HD																					
3.0mm HD Bone Tap	TST2030HD																					
2.8mm HD Drill	TSD2028HD																					
2.8mm HD Drill, Extended Shank	TSD4028HD																					
3.4mm HD Crestal Bone Drill	TSC2034HD																					
3.4mm HD Bone Tap	TST2034HD																					
3.2mm HD Drill	TSD2032HD																					
3.2mm HD Drill, Extended Shank	TSD4032HD																					
3.8mm HD Crestal Bone Drill	TSC2038HD																					
3.8mm HD Bone Tap	TST2038HD																					
3.7mm HD Drill	TSD2037HD																					
3.7mm HD Drill, Extended Shank	TSD4037HD																					
4.1mm HD Drill	TSD2041HD																					
4.1mm HD Drill, Extended Shank	TSD4041HD																					
4.6mm HD Crestal Bone Drill	TSC2046HD																					
4.6mm HD Bone Tap	TST2046HD																					
4.7mm HD Drill	TSD2047HD																					
4.7mm HD Drill, Extended Shank	TSD4047HD																					
5.4mm HD Drill	TSD2054HD																					
5.4mm HD Drill, Extended Shank	TSD4054HD																					
5.8mm HD Crestal Bone Drill	TSC2058HD																					
5.8mm HD Bone Tap	TST2058HD																					

Important Note:
All drills should be replaced after they have been used 12 to 20 times, depending upon bone density.

Shop on-line at www.biohorizons.com