

Tapered IM Drill Usage Chart

Part Description	Part Number	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Tapered Length Drill, 7.0mm x 7.5mm	TIMD7007																				
Tapered Length Drill, 8.0mm x 7.5mm	TIMD8007																				
Tapered Length Drill, 7.0mm x 9.0mm	TIMD7009																				
Tapered Length Drill, 8.0mm x 9.0mm	TIMD8009																				
Tapered Length Drill, 7.0mm x 10.5mm	TIMD7010																				
Tapered Length Drill, 8.0mm x 10.5mm	TIMD8010																				



TIMD7007 TIMD8007 TIMD7009 TIMD8009 TIMD7010 TIMD8010

Important Note:
All drills should be replaced after they have been used 12 to 20 times, depending upon bone density.